Clinic Freezer Temperature Logger for Pfizer Vaccine Management

- Pfizer vaccine can be transferred from Ultra Cold Freezers operated by the RFDS into clinic freezers kept between -15 and -25◦C
- Pfizer vaccine can be stored between -15 and -25 ◦C for up to two weeks, at which time they must be transferred to clinic vaccine refrigerator to defrost

1. Monitor temperatures of freezer closely.
2. Write your initials below in "Staff Initials", and note the time in "Exact Time".
3. Document current temperature twice each workday.
4. Document min/max temps once each workday i.e. in the morning.
5. Put an "X" in the row that corresponds to the freezer's temperature.
6. If any out-of-range temp observed (**too warm (above -15◦ C) or too cold (below -25◦ C)**), alert clinic manager immediately. Clinic manager to discuss with medical director and Australian Government Vaccine Operations Centre (1800 318 208).

Clinic Name: _____ Year: _____

Date		/		/		/		/		/		/		/		/		/		/		/		/	
Staff Initials																									
Exact Time		AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM
Min/Max Temp (◦C)																									
ACCEPTABLE TEMPERATURE RANGE: -15◦C to -25◦C | Write any out-of-range temps below and alert clinic manager immediately!																									
ACCEPTABLE TEMPERATURES	-15																								
	-16																								
	-17																								
	-18																								
	-19																								
	-20																								
	-21																								
	-22																								
	-23																								
	-24																								
	-25																								
ACTION	Write any out-of-range temps here																								